

Declaration of Conformity

According to ISO/IEC 17050-1 and EN 17050-1



Manufacturer's Name: Technologies HumanWare Inc.
Manufacturer's Address: 1800, rue Michaud, Drummondville, Québec, CANADA, J2C 7G7

Declares, that the products:

Product Names: Victor Reader Stratus 4, Victor Reader Stratus 12, Victor Reader Stratus 4M,
Victor Reader Stratus 12M
Model Numbers: 402VR, 403VR, 412VR, 413VR

Conforms to the following Product Specifications and Regulations:

Safety IEC 60601:1988 +A1:1991 +A2:1995
Medical electrical equipment –
Part 1: General requirements for safety and essential performance.

IEC 60950-1:2001
Information technology equipment. Safety. General requirements

EMC EN 60601-1-2: 2001 + A1:2006
Medical electrical equipment – General requirements for safety –
Collateral standard: Electromagnetic compatibility. - Requirements and tests

EN55024:1998 + A1: 2001 + A2: 2003
Information technology equipment – Immunity characteristics

EN 55022: 2006 + A1: 2007
Information technology equipment – Radio disturbance characteristics

Technical aids for disabled persons EN 12182:1999

Application of risk management to medical devices ISO 14971:2000

Supplementary Information:

1. This product complies with the requirements of the Low Voltage Directive 2006/95/EC, the EMC Directive 2004/108/EC and the Medical Device Directive 93/42/EEC, and carries the CE-marking accordingly.
2. Based on information derived from supplier material declarations and sub-contractor assembly certifications, we declare that the product(s) listed above complies with the RoHS Directive 2002/95/EC. In addition, it complies with the WEEE Directive 2002/96/EC.
3. This Device complies with Part 15 of the FCC Rules. Operation is subject to the following two Conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.
4. The product was tested in a typical configuration.

Date: June 15, 2011


Mr. Pierre Hamel
Vice-president Research & Development
HumanWare Canada

Declaration of Conformity

According to ISO/IEC 17050-1 and EN 17050-1

Humanware™

Manufacturer's Name: Technologies HumanWare, Inc.
Manufacturer's Address: 1800, Michaud, Drummondville, Quebec, Canada, J2C 7G7

Declares, that the products:

Product Names: Victor Reader Stream

Model Numbers: 503 VRC

Conforms to the following Product Specifications and Regulations:

Safety EN 60950-1:2006 +A11:2009 +A1:2010 +A12:2011

EMC EN 55024: 2010
EN 55022: 2010
CISPR 22: Edition 6.0 2008-09
EN 301 489-17 V2.1.1 (2009-05)
EN 300 328 V1.7.1 (2006-10)
EN 61000-3-2: 2006 + A1: 2009 + A2: 2009
EN 61000-3-3: 2008
EN50371: 2002

Technical aids for disabled persons EN 12182:1999

Supplementary Information:

1. This device also complies with FCC 47 CFR Part 15, Subpart B, Subpart C, §15.247 (USA), RSS-210, Issue 8 Annex 8, ICES-003 Issue 5 August 2012 (Canada), AS/NZS CISPR 22: 2009 (International)
2. Based on information derived from supplier material declarations and sub-contractor assembly certifications, we declare that the product(s) listed above complies with the RoHS Directive 2002/95/EC. In addition, it complies with the WEEE Directive 2002/96/EC.

Date: October 22, 2012



Mr. Pierre Hamel
Vice-president Research & Development
Technologies Humanware Inc