

EC DECLARATION OF CONFORMITY

Technologies HumanWare Inc. (1800 rue Michaud, Drummondville, Quebec, Canada, J2C 7G7) declare that the following medical devices:

Prodigi CCTV all models: Duo, Desktop 20, 24; Duo, Desktop XY 20, 24; Tablet/ Touch & Connect

Explore, all models: Explore 5, 7

SmartView / MyView: all models

Brailiant Braille displays all models: BI14, BI32, BI40, B40, B80

Victor Reader Stream Stratus Daisy player all models: Stratus Basic 4, 4M, 4H & 12 H

Victor Reader Stream Daisy Player all models

Braille Note Touch all models

Are in conformity with the following standards:

- CE marked and compliant with 93/42 EEC directive for medical devices
- EN ISO 14971:2009 Medical devices – Application of risk management to medical devices
- EN 12182:2012 Technical Aids for disabled persons – General requirements and test methods
- EN 12182 Electric Safety and Electromagnetic compatibility

Date of issue 14/06/2016

Place of issue: Montreal, Quebec, Canada

Name of signatory: 

Name of Signatory: Pedro S Polson

Position of signatory Chief Executive Officer, HumanWare Europe

Humanware[™]

Russell Smith House
2 Bullmatt Business Centre
Northampton Road Rushden
Northamptonshire NN10 6AR